

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

OCT 05 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130579</u> ✓	
A	Name <u>Michael Bussey</u>	ID# <u>0111</u>	Date <u>7/26/2009</u> ✓
	Agency <u>APD</u>	Phone # <u>786-8500</u>	
	Instrument Location <u>Anchorage</u> ✓		
B	Alco S/N <u>X172946</u>	Target Value <u>.084</u> ✓	High Pressure <u>250</u>
	Alco Test Value Average <u>.083</u> ✓	<u>.084</u> ✓	
	1 st Alco	2 nd Alco	
Signature <u>Michael Bussey</u> ✓			
(OVER)			

023
10/6/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not Write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 6th day of Oct., 2009.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130579✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579 ✓

JULY 26, 2009

OPERATOR'S NAME:

BUSEY, M

OPERATOR'S NUMBER: 0111

SUBJECT'S LAST NAME:

NON-DRINKER

SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 000000

DEPT/AGENCY: ANCL

CASE/REPORT: 09-00000

TEST TYPE: V

ALCO TARGET VALUE: .084 ✓

ALCO S/N: X172946 ✓

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.14 in

ALCO TARGET

BLANK TEST

INTERNAL STANDARD

ALCO TV 30.14 in

BLANK TEST

SUBJECT SAMPLE

BLANK TEST

ALCO TV 30.16 in

BLANK TEST

.084

.000

VERIFIED

.083 ✓

.000

.000

.000

.084 ✓

.000

250 lb. High Pressure
Busey 0111 7/26/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579 ✓

JULY 26, 2009

TIME 00:56

--- DIAGNOSTIC CHECK ---

COMPUTER:

OKAY

PROGRAM:

OKAY

SOFTWARE DATE:

12/11/08

HEATERS

SAMPLE CHAMBER:

49c

BREATH TUBE:

40c

BAROMETER:

30.14 in

FLOW DETECTOR:

OKAY

PUMP

HIGH SPEED:

OKAY

DETECTOR:

OKAY

FILTERS:

OKAY

QUARTZ STANDARD:

OKAY

CALIBRATION:

OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~^